

SAN FRANCISCO BAY NAVAL SHIPYARD  
YARD ROUTE SLIP

CODE		NAME		EXTENSION	DATE
967 Radiac Facility		George Menize		2774/2776	9/7/73
TO CODE	INITIALED		REMARKS		
	BY	DATE			
833			Bill;		
Baldassari			Radiac set AN/PDR-18B ser # 1043 plant acct# 217-200965 which is part of the Disaster Control Inventory, has a leaky source and is being reported monthly on the attached report.		
630.3					
Kearney			I would like to suggest we work toward removing this item from the Shipyard soon rather than later because we still have a few people who can remove the source (Sr 90 300 uci) and a disposal container in bldg 146 to place it in.		
950.2					
Nelson			I can arrange to have the source removed and disposed of, however I would like to know the disposition of the radiac prior to doing so. I would <del>like</del> suggest you discuss this with Lee Kearney. It may be the unit should be disposed of as scrap (after source removal) or if needed, perhaps the radiac		

