5/N-0104-[illegible]762 (REV. 11-67)

DEPARTMENT OF THE NAVY

Memorandum

DATE: 6 Dec 1974

FROM : L. P. McNeice, Code 034

TO : L. Kerney, Code 630.3, San Francisco, CA 94135

SUBJ : Radioactive Survey; results of

1. As requested, the following report is submitted on wipe tests performed at the locations indicated. In addition, surveys made with both an AN/PDR-56 and an E-140N at the same locations, indicated no detectable activity.

(a)	[redacted]	N.D.A.
(b)	Building 146	N.D.A.
(c)	Building 214, RM 105	N.D.A.
(d)	Building 253, 6th Floor	N.D.A.
(e)	Building 351A	N.D.A.

L. P. McNEICE

ENCLOSURE (2)

☆ U.S. GOVERNMENT PRINTING OFFICE: 1970-977-399-147

Bldg [illegible]

Radiographers Vault & Waste Disposal Storage

Note: [illegible] for Waste Disposal Storage [illegible] 10/73 to 6/74



Scale 1" = 1 foot

29 WIPES
2 WIPE ON [illegible]

Construction: Concrete (Walls & Floor)

ENCLOSURE (3)

Bldg 146 (Ground Floor) 8/21/74

Radioactive Waste Disposal Drum Storage

[illegible] 10/53.

Drum Location



Scale 1 square = 1 ft.

Construction: Concrete Walls & Floor

ENCLOSURE (3)

Bldg 214 Rm 105 8/15/74

Health Physics Counting Rm.



Scale 1 Square = 1 foot

Construction: Wood frame

ENCLOSURE (3)



Construction: Concrete Walls & Floor

Measurements made 8/21/74

CALIBRATION ROOM

BLDG. 253 6TH FLOOR

SFNS SKETCH NO. 10555
CODE 288 3/20/58

Bldg 35 A 8/15/74

Radioc Repair Facility



Scale: 1 Square = 3 ft

Construction: Wood Frame

ENCLOSURE (3)